

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/29/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:43
INTERNAL STANDARD	VERIFIED	15:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:46
BLANK TEST	0.00	15:47

PO Kenneth W Smith #301

7/29/19

