

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/29/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:34
INTERNAL STANDARD	VERIFIED	08:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:38

P.D. K W Smith 301

3/29/19

