

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/02/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:01
BLANK TEST	0.00	10:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:03
BLANK TEST	0.00	10:04

Handwritten signature/initials

