

SUPERVISOR TEST



WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/17/2019

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:24
INTERNAL STANDARD	VERIFIED	02:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28



*Charles B. Geisler* #215