

SUPERVISOR TEST




-----  
 WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 05/08/2019

OPERATOR NAME:  
 CHARLES B GEISLER  
 Permit #: 33757  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 19030

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 23:30 |
| INTERNAL STANDARD       | VERIFIED | 23:30 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 23:30 |
| BLANK TEST              | 0.00     | 23:31 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 23:31 |
| BLANK TEST              | 0.00     | 23:32 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 23:33 |
| BLANK TEST              | 0.00     | 23:33 |



*Charles B. Geisler*