

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/22/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:08
INTERNAL STANDARD	VERIFIED	10:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:09
BLANK TEST	0.00	10:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:10
BLANK TEST	0.00	10:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:11
BLANK TEST	0.00	10:12

8070/104

