

SOLUTION CHANGE



WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/01/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:03
INTERNAL STANDARD	VERIFIED	03:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:06
BLANK TEST	0.00	03:07



Charles B. Geisler #224