

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/08/2019

OPERATOR NAME:  
TYLER J MORSE  
Permit #: 41959  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:39
INTERNAL STANDARD	VERIFIED	13:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:40
BLANK TEST	0.00	13:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:41
BLANK TEST	0.00	13:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:42
BLANK TEST	0.00	13:43

