

SOLUTION CHANGE




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/03/2019

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:10
INTERNAL STANDARD	VERIFIED	03:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:14



*Handwritten signature: Charles B. Geisler*