

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/27/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:31
INTERNAL STANDARD	VERIFIED	04:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:32
BLANK TEST	0.00	04:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:33
BLANK TEST	0.00	04:34

