

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/01/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:30
INTERNAL STANDARD	VERIFIED	16:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:31
BLANK TEST	0.00	16:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:32
BLANK TEST	0.00	16:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:33
BLANK TEST	0.00	16:34

