

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/22/2019

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:29
INTERNAL STANDARD	VERIFIED	09:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:32
BLANK TEST	0.00	09:32

CO. # 301

11/22/19

