

SUPERVISOR TEST


WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/06/2019

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:10
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12



Charles B Geisler #248