

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/28/2019

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:46
INTERNAL STANDARD	VERIFIED	08:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:48
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:49
BLANK TEST	0.00	08:50

RS/S

