

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/09/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:20
INTERNAL STANDARD	VERIFIED	16:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:20
BLANK TEST	0.00	16:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:21
BLANK TEST	0.00	16:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:23
BLANK TEST	0.00	16:23

Charles B. Geisler
#24

