

SUPERVISOR TEST

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 04/07/2020

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:32
INTERNAL STANDARD	VERIFIED	13:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:34
BLANK TEST	0.00	13:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:35
BLANK TEST	0.00	13:36

Signature: Sgt. M. B. [unclear] #246

