

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/14/2020

OPERATOR NAME: *Do Carpenter #296*
GEORGE T CARPENTER
Permit #: 38610
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:22
INTERNAL STANDARD	VERIFIED	09:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26

