

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/28/2020

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:07
INTERNAL STANDARD	VERIFIED	11:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:07
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:08
BLANK TEST	0.00	11:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:09
BLANK TEST	0.00	11:10



John D Bateholts