

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/26/2020

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:14
INTERNAL STANDARD	VERIFIED	12:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:15
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:17
BLANK TEST	0.00	12:18

Chester town

Don J Long *844*

