

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/17/2020

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:04
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:05
BLANK TEST	0.00	12:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:07
BLANK TEST	0.00	12:08

P.O. Kenneth W Smith #301

11/17/2020

