

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/14/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:45
INTERNAL STANDARD	VERIFIED	15:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:47
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:49

Handwritten signature and initials
#2411

