

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/12/2020

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:46
INTERNAL STANDARD	VERIFIED	10:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:47
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:48
BLANK TEST	0.00	10:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:49
BLANK TEST	0.00	10:50



P.O. *[Signature]* #714