

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/02/2020

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:38
INTERNAL STANDARD	VERIFIED	06:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:39
BLANK TEST	0.00	06:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:40
BLANK TEST	0.00	06:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:41
BLANK TEST	0.00	06:42



*PO Michael Campbell #757*