

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/26/2020

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:23
INTERNAL STANDARD	VERIFIED	13:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:23
BLANK TEST	0.00	13:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:24
BLANK TEST	0.00	13:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:26
BLANK TEST	0.00	13:27

Qby
pc. 1 *Don* *8/26/20*

