

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/14/2020

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:49
INTERNAL STANDARD	VERIFIED	14:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:51
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:52
BLANK TEST	0.00	14:53



Handwritten signature and date: P. Stua 10/14/20