

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/17/2020

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:02
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:05
BLANK TEST	0.00	13:06

P.O. *[Signature]* #301

11/17/2020

