

SUPERVISOR TEST


WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/06/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:55
INTERNAL STANDARD	VERIFIED	23:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:56
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:57
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:59



Charles B. Geisler *Hend*