

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/06/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:08
INTERNAL STANDARD	VERIFIED	14:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:09
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12



Sgt. A. B. Geisler #245