

SUPERVISOR TEST




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/28/2020

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:02
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:03
BLANK TEST	0.00	14:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:04
BLANK TEST	0.00	14:05



*Charles B. Geisler*