

SUPERVISOR TEST




WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/23/2020

OPERATOR NAME:
 CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:22
INTERNAL STANDARD	VERIFIED	14:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:24
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:25
BLANK TEST	0.00	14:26



Signature: Sgt. Charles B. Geisler # 2018