

SUPERVISOR TEST



WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/29/2020

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:15
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:16
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:17
BLANK TEST	0.00	13:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:18
BLANK TEST	0.00	13:19



*Charles B. Geisler* #2416

SUPERVISOR TEST

-----  
WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506  
-----

Date: 03/29/2020

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:12
INTERNAL STANDARD	VERIFIED	13:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:13

Communication Error 1

