

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/06/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:52
INTERNAL STANDARD	VERIFIED	12:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56

[Handwritten signatures and initials]

