

SOLUTION CHANGE




WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/02/2020

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:59
INTERNAL STANDARD	VERIFIED	05:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:01
BLANK TEST	0.00	06:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:02
BLANK TEST	0.00	06:03



*PO [Signature] #157*