

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 06/07/2020

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:36
INTERNAL STANDARD	VERIFIED	13:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:38
BLANK TEST	0.00	13:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:40

