

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/07/2020

OPERATOR NAME:
TYLER J MORSE
Permit #: 41959
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:26
INTERNAL STANDARD	VERIFIED	17:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:26
BLANK TEST	0.00	17:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:28
BLANK TEST	0.00	17:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:29
BLANK TEST	0.00	17:30



(Signature) #447 *TYLER J MORSE*