

SUPERVISOR TEST

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 07/14/2020

OPERATOR NAME: *PO Carpenter #296*
 GEORGE T CARPENTER
 Permit #: 38610
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:43
INTERNAL STANDARD	VERIFIED	09:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:43
BLANK TEST	0.00	09:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:46
BLANK TEST	0.00	09:47

