

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/28/2020

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:44
INTERNAL STANDARD	VERIFIED	10:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:45
BLANK TEST	0.00	10:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:46
BLANK TEST	0.00	10:47



John D. Bateholts 07/28/2020