

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/01/2020

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:49
INTERNAL STANDARD	VERIFIED	13:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:50
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:51
BLANK TEST	0.00	13:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:53



Sgt. C.B. Geisler #241