

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/26/2020

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:53
INTERNAL STANDARD	VERIFIED	12:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:55
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:56
BLANK TEST	0.00	12:57

Wernersberg
P.O. Donnie J Long 844

