

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/17/2020

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:31
INTERNAL STANDARD	VERIFIED	12:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:32
BLANK TEST	0.00	12:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35

PO [Signature] 11/30/20

11/17/2020

