

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/15/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:50
INTERNAL STANDARD	VERIFIED	13:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:51
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54

Charles B Geisler #201

