

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/22/2021

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:19
INTERNAL STANDARD	VERIFIED	14:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:20
BLANK TEST	0.00	14:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:21
BLANK TEST	0.00	14:22

Josh M Lopez

