

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/31/2021

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:35
INTERNAL STANDARD	VERIFIED	16:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:35
BLANK TEST	0.00	16:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:37
BLANK TEST	0.00	16:38



Signature ST