

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/22/2021

OPERATOR NAME:  
KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:52
INTERNAL STANDARD	VERIFIED	10:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:55
BLANK TEST	0.00	10:56

P.O. *[Signature]* #301

03/22/2021

