

SOLUTION CHANGE

 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 05/01/2021

OPERATOR NAME:
 RYAN C SCHROECK
 Permit #: 41316
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:39
INTERNAL STANDARD	VERIFIED	23:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43

RS/S

