

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/08/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:48
INTERNAL STANDARD	VERIFIED	11:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52

Ch. B. Geisler

