

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/06/2021

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:54
BLANK TEST	0.00	09:55

Charles B. Geisler #241

