

SOLUTION CHANGE

WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 07/02/2021

OPERATOR NAME:
 KENNETH W SMITH
 Permit #: 35604
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:05
INTERNAL STANDARD	VERIFIED	15:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:09

P.O. Smith #301

7/2/2021

