

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/16/2021

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:35
INTERNAL STANDARD	VERIFIED	21:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:38
BLANK TEST	0.00	21:39

RS/S

