

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/08/2021

OPERATOR NAME:
RYAN C SCHROECK
Permit #: 41316
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:49
INTERNAL STANDARD	VERIFIED	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:51
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:52
BLANK TEST	0.00	23:53

907/1724

